
Midmark M11 - v1.0.0
Total Cycles: 4130

Name: RITTER 1
Sterilizer ID: V2178198
12/18/2023 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.5 F 0.0
2:00 174.1 F 3.2
4:00 178.5 F 6.9
6:00 224.1 F 11.9
8:00 223.7 F 5.7
10:00 234.0 F 10.4
12:00 231.7 F 7.4
14:00 243.1 F 12.5
16:00 240.2 F 10.8
18:00 253.0 F 17.1
20:00 264.2 F 23.6
22:00 270.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 271.9 F 28.8
2:00 272.1 F 28.8
3:00 272.6 F 29.1
4:00 272.6 F 29.3

Min 271.6 F 28.7
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 208.8 F 0.0
10:00 248.5 F 0.0
15:00 233.8 F 0.0
20:00 224.1 F 0.0
25:00 231.0 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:21
CYCLE COMPLETE

12/18/2023 12:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____