
Midmark M11 - v1.0.5
Total Cycles: 4010

Name: RITTER2
Sterilizer ID: V2313228
10/08/2024 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.6 F 0.0
2:00 102.4 F 0.0
4:00 146.7 F 0.0
6:00 152.0 F 1.1
8:00 166.9 F 3.9
10:00 185.2 F 6.9
12:00 200.6 F 10.3
14:00 212.6 F 8.4
16:00 224.0 F 7.2
18:00 235.7 F 11.7
20:00 234.0 F 9.0
22:00 240.0 F 10.7
24:00 245.7 F 13.6
26:00 257.1 F 19.7
28:00 267.3 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.6 F 29.7
2:00 271.4 F 29.0
3:00 272.4 F 29.8
4:00 271.8 F 29.2

Min 271.1 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 196.4 F 0.0
10:00 217.4 F 0.0
15:00 221.6 F 0.0
20:00 225.1 F 0.0
25:00 232.0 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:31
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:08:08
CYCLE COMPLETE

10/08/2024 12:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____