
Midmark M11 - v1.0.0
Total Cycles: 743

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2021 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 181.9 F 3.0
4:00 205.9 F 7.9
6:00 225.5 F 8.9
8:00 227.4 F 9.7
10:00 234.7 F 8.3
12:00 244.8 F 12.5
14:00 245.0 F 12.5
16:00 258.3 F 19.6
18:00 269.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 271.5 F 28.7
2:00 271.7 F 28.7
3:00 271.8 F 28.7
4:00 271.8 F 28.6

Min 271.4 F 28.6
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.6 F 0.0
5:00 241.5 F 0.0
10:00 235.9 F 0.0
15:00 226.2 F 0.0
20:00 227.3 F 0.0
25:00 227.6 F 0.0
30:00 221.2 F 0.0

Drying Complete

FILLING: 1:58
HEATING: 18:53
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 00:57:45
CYCLE COMPLETE

03/24/2021 02:26 PM
ACCEPTED: X
REJECTED:
Operator: U01