
Midmark M11 - v1.0.5
Total Cycles: 1265

Name: RITTER2
Sterilizer ID: V2313228
07/07/2022 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.6 F 0.0
2:00 104.5 F 0.5
4:00 117.2 F 2.5
6:00 145.6 F 5.6
8:00 174.3 F 9.5
10:00 194.5 F 11.1
12:00 215.2 F 6.8
14:00 224.2 F 11.2
16:00 227.8 F 7.1
18:00 241.3 F 12.5
20:00 238.7 F 10.2
22:00 251.0 F 16.2
24:00 261.5 F 22.3
26:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.9 F 29.7
2:00 271.7 F 29.3
3:00 272.4 F 29.8
4:00 271.8 F 29.1

Min 271.6 F 29.1
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 238.1 F 0.0
10:00 224.8 F 0.0
15:00 231.8 F 0.0
20:00 234.4 F 0.0
25:00 231.5 F 0.0
30:00 213.8 F 0.0

Drying Complete

FILLING: 1:02
HEATING: 27:33
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:05:42
CYCLE COMPLETE

07/07/2022 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____