
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003886
10:32 27/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 2:59
STERILIZING 4:20
136.5 C 222kPa 4:20
136.2 C 222kPa 4:50
136.4 C 223kPa 5:20
136.4 C 224kPa 5:50
136.4 C 224kPa 6:20
136.4 C 223kPa 6:50
136.3 C 223kPa 7:20
136.4 C 224kPa 7:50
136.4 C 224kPa 7:50
VENTING 7:50
AIR DRYING 9:15
CYCLE COMPLETE 10:10

Digital Signature #
7FF6E8157B177A5C
