
Midmark M11 - v1.0.5
Total Cycles: 1414

Name: RITTER2
Sterilizer ID: V2313228
08/25/2022 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.2 F 0.0
2:00 195.6 F 2.5
4:00 202.8 F 7.2
6:00 223.2 F 12.8
8:00 228.6 F 8.8
10:00 238.7 F 10.6
12:00 242.8 F 12.8
14:00 242.9 F 12.1
16:00 255.9 F 19.0
18:00 267.4 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.4 F 29.5
2:00 271.6 F 28.8
3:00 272.4 F 29.6
4:00 271.9 F 28.9

Min 271.5 F 28.7
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 204.7 F 0.0
10:00 211.4 F 0.0
15:00 217.1 F 0.0
20:00 218.4 F 0.0
25:00 217.7 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

08/25/2022 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____