
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002237
9:13 11/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:02
STERILIZING 3:05
136.4 C 224kPa 3:05
136.2 C 223kPa 3:35
136.3 C 224kPa 4:05
136.4 C 223kPa 4:35
136.4 C 223kPa 5:05
136.0 C 221kPa 5:35
136.2 C 224kPa 6:05
136.4 C 222kPa 6:35
136.4 C 222kPa 6:35
VENTING 6:35
AIR DRYING 7:28
CYCLE COMPLETE 37:28

Digital Signature #
D23A58842A665A2A
