
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002630
8:55 28/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:58
STERILIZING 3:06
136.4 C 224kPa 3:06
136.2 C 225kPa 3:36
136.4 C 226kPa 4:06
136.4 C 225kPa 4:36
136.4 C 226kPa 5:06
136.3 C 224kPa 5:36
136.3 C 225kPa 6:06
136.4 C 222kPa 6:36
136.4 C 222kPa 6:36
VENTING 6:36
AIR DRYING 7:33
CYCLE COMPLETE 8:19

Digital Signature #
9CE02B0A2110C5B9
