
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 000312
12:34 20/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:09
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:09

Digital Signature #
45854F0E42B426FE
