

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002032  
9:20 13/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:07  
STERILIZING 3:12  
136.5 C 224kPa 3:12  
136.3 C 224kPa 3:42  
136.4 C 225kPa 4:12  
136.4 C 225kPa 4:42  
136.3 C 225kPa 5:12  
136.4 C 225kPa 5:42  
136.4 C 226kPa 6:12  
136.2 C 224kPa 6:42  
136.2 C 224kPa 6:42  
VENTING 6:42  
AIR DRYING 7:33  
CYCLE COMPLETE 19:30  
-----

Digital Signature #  
F384502C78E5B0AD  
-----