
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000078
10:53 14/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:23
STERILIZING 4:43
136.4 C 226kPa 4:43
136.4 C 225kPa 5:13
136.0 C 222kPa 5:43
136.4 C 225kPa 6:13
136.4 C 222kPa 6:43
136.1 C 222kPa 7:13
136.3 C 223kPa 7:43
136.3 C 224kPa 8:13
136.3 C 224kPa 8:13
VENTING 8:13
AIR DRYING 9:09
CYCLE COMPLETE 15:18

Digital Signature #
09602259FC3A5971
