

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 005705  
15:02 07/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:31  
STERILIZING 3:34  
136.5 C 221kPa 3:34  
136.4 C 222kPa 4:05  
136.4 C 221kPa 4:35  
136.4 C 225kPa 5:05  
136.4 C 221kPa 5:35  
136.2 C 221kPa 6:05  
136.1 C 218kPa 6:35  
136.4 C 223kPa 7:05  
136.4 C 223kPa 7:05  
VENTING 7:05  
AIR DRYING 7:52  
CYCLE COMPLETE 9:30

-----  
Digital Signature #  
AFA66D90CA18FC92  
-----