
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001527
12:07 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:51
STERILIZING 3:52
136.4 C 224kPa 3:52
136.2 C 223kPa 4:22
136.4 C 227kPa 4:52
136.3 C 225kPa 5:22
136.1 C 222kPa 5:52
136.3 C 223kPa 6:22
136.3 C 224kPa 6:52
136.1 C 224kPa 7:22
136.1 C 224kPa 7:22
VENTING 7:22
AIR DRYING 8:19
CASSETTE REMOVED
CYCLE COMPLETE 27:14

Digital Signature #
09B8306FC6947293
