
Midmark M11 - v1.0.0
Total Cycles: 3274

Name: RITTER 1
Sterilizer ID: V2178198
05/24/2023 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.8 F 0.0
2:00 171.5 F 0.0
4:00 186.5 F 2.5
6:00 204.1 F 7.5
8:00 220.8 F 12.7
10:00 227.5 F 8.2
12:00 240.3 F 12.5
14:00 241.4 F 11.6
16:00 239.8 F 10.2
18:00 253.9 F 17.2
20:00 265.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.3 F 28.7
2:00 271.4 F 28.2
3:00 272.4 F 28.8
4:00 271.4 F 28.1

Min 271.3 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 233.4 F 0.0
10:00 232.4 F 0.0
15:00 229.4 F 0.0
20:00 230.7 F 0.0
25:00 228.4 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:51
CYCLE COMPLETE

05/24/2023 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____