
Midmark M11 - v1.0.5
Total Cycles: 1516

Name: RITTER2
Sterilizer ID: V2313228
09/21/2022 02:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.0 F 0.0
2:00 202.4 F 4.2
4:00 205.4 F 9.5
6:00 221.8 F 7.4
8:00 234.3 F 11.2
10:00 234.1 F 9.1
12:00 238.5 F 9.6
14:00 248.3 F 15.1
16:00 260.7 F 22.1
18:00 271.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.5 F 30.0
2:00 271.5 F 29.3
3:00 271.8 F 30.1
4:00 272.4 F 29.8

Min 271.1 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 232.7 F 0.0
10:00 229.4 F 0.0
15:00 232.4 F 0.0
20:00 226.7 F 0.0
25:00 232.8 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 18:35
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

09/21/2022 03:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____