

-----  
Midmark M11 - v1.0.5  
Total Cycles: 835

Name: RITTER2  
Sterilizer ID: V2313228  
02/24/2022 10:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.8 F 0.0  
2:00 151.7 F 2.6  
4:00 166.9 F 6.1  
6:00 190.0 F 11.0  
8:00 220.0 F 4.6  
10:00 229.6 F 9.6  
12:00 235.4 F 9.0  
14:00 242.7 F 12.5  
16:00 242.3 F 11.5  
18:00 255.6 F 18.4  
20:00 267.7 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.4 F 29.0  
2:00 271.6 F 28.8  
3:00 272.6 F 29.0  
4:00 271.3 F 28.5  
  
Min 271.3 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.8 F 0.0  
5:00 225.8 F 0.0  
10:00 224.5 F 0.0  
15:00 221.7 F 0.0  
20:00 219.2 F 0.0  
25:00 217.8 F 0.0  
30:00 222.3 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 21:19  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:41  
CYCLE COMPLETE

02/24/2022 11:41 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02