
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000826
14:22 17/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:06
STERILIZING 4:12
136.0 C 228kPa 4:12
136.4 C 233kPa 4:42
136.2 C 230kPa 5:12
136.3 C 232kPa 5:42
136.3 C 231kPa 6:12
136.1 C 231kPa 6:42
136.4 C 232kPa 7:12
136.4 C 232kPa 7:42
136.4 C 232kPa 7:42
VENTING 7:43
AIR DRYING 8:35
CYCLE COMPLETE 23:35

Digital Signature #
E329B21BF6457A0B
