
Midmark M11 - v1.0.0
Total Cycles: 4600

Name: RITTER 1
Sterilizer ID: V2178198
06/03/2024 07:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 71.8 F 0.3
4:00 89.3 F 1.3
6:00 118.4 F 3.3
8:00 143.5 F 6.0
10:00 165.8 F 9.3
12:00 186.1 F 12.9
14:00 206.6 F 3.8
16:00 221.1 F 6.9
18:00 229.0 F 10.2
20:00 234.4 F 9.7
22:00 235.6 F 9.6
24:00 240.9 F 10.2
26:00 243.3 F 12.0
28:00 254.4 F 17.6
30:00 264.1 F 23.4
32:00 272.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 271.7 F 28.3
2:00 272.0 F 28.5
3:00 272.3 F 28.8
4:00 272.4 F 29.1

Min 271.4 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 210.0 F 0.0
10:00 215.2 F 0.0
15:00 229.7 F 0.0
20:00 229.9 F 0.0
25:00 227.2 F 0.0
30:00 225.3 F 0.0

Drying Complete

FILLING: 0:36
HEATING: 32:09
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:07
CYCLE COMPLETE

06/03/2024 08:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____