
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000736
8:03 23/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:06
STERILIZING 3:09
136.4 C 220kPa 3:09
136.3 C 232kPa 3:39
136.4 C 230kPa 4:09
136.1 C 229kPa 4:39
136.4 C 232kPa 5:09
136.4 C 232kPa 5:39
136.3 C 232kPa 6:09
136.1 C 232kPa 6:39
136.1 C 232kPa 6:39
VENTING 6:39
AIR DRYING 7:29
CYCLE COMPLETE 7:37

Digital Signature #
9808979CDECE2702
