

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1796

Name: RITTER 1  
Sterilizer ID: V2178198  
04/05/2022 12:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.5 F 0.0  
2:00 186.2 F 4.1  
4:00 192.7 F 9.0  
6:00 221.4 F 7.1  
8:00 227.4 F 8.9  
10:00 238.7 F 10.2  
12:00 242.4 F 11.9  
14:00 240.5 F 10.5  
16:00 254.3 F 17.4  
18:00 266.0 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.0 F 28.4  
2:00 272.7 F 28.9  
3:00 271.8 F 28.4  
4:00 271.8 F 28.5  
  
Min 271.4 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
5:00 233.7 F 0.0  
10:00 216.7 F 0.0  
15:00 200.6 F 0.0  
20:00 216.8 F 0.0  
25:00 236.9 F 0.0  
30:00 227.3 F 0.0  
  
Drying Complete

FILLING: 1:32  
HEATING: 19:35  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:15  
CYCLE COMPLETE

04/05/2022 01:32 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_