
Midmark M11 - v1.0.0
Total Cycles: 2009

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2022 11:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.3 F 0.0
2:00 146.5 F 2.6
4:00 167.9 F 5.9
6:00 191.1 F 10.3
8:00 211.2 F 5.0
10:00 221.7 F 8.4
12:00 234.3 F 13.2
14:00 234.7 F 9.3
16:00 237.8 F 9.4
18:00 245.8 F 13.2
20:00 257.6 F 19.6
22:00 268.6 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.2
2:00 271.7 F 28.5
3:00 271.6 F 28.8
4:00 272.3 F 29.1

Min 271.2 F 28.2
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 206.7 F 0.0
10:00 221.8 F 0.0
15:00 231.4 F 0.0
20:00 233.4 F 0.0
25:00 227.9 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:15
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:20
CYCLE COMPLETE

06/08/2022 12:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____