
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000762
7:15 28/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:06
STERILIZING 3:16
136.0 C 228kPa 3:16
136.4 C 233kPa 3:46
136.4 C 230kPa 4:16
136.4 C 232kPa 4:46
136.3 C 230kPa 5:16
136.4 C 232kPa 5:46
136.4 C 233kPa 6:16
136.2 C 230kPa 6:46
136.2 C 230kPa 6:46
VENTING 6:46
AIR DRYING 7:41
CYCLE COMPLETE 7:43

Digital Signature #
B510837FC27B29BD
