
Midmark M11 - v1.0.0
Total Cycles: 1790

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2022 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.9 F 0.0
2:00 139.5 F 2.4
4:00 161.2 F 6.1
6:00 185.7 F 10.5
8:00 211.2 F 4.1
10:00 226.2 F 8.5
12:00 236.2 F 13.5
14:00 237.7 F 10.0
16:00 236.9 F 8.5
18:00 248.6 F 14.5
20:00 260.8 F 21.3
22:00 271.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.1 F 28.7
2:00 272.1 F 28.9
3:00 271.9 F 28.3
4:00 272.4 F 29.1

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 229.9 F 0.0
10:00 212.7 F 0.0
15:00 202.4 F 0.0
20:00 192.0 F 0.0
25:00 211.8 F 0.0
30:00 233.1 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:31
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

04/04/2022 02:35 PM
ACCEPTED: X
REJECTED:
Operator: U02