
Midmark M11 - v1.0.0
Total Cycles: 2907

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2023 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.5 F 0.0
2:00 208.1 F 4.9
4:00 205.2 F 10.0
6:00 217.9 F 6.1
8:00 232.6 F 10.6
10:00 234.4 F 7.3
12:00 242.9 F 12.1
14:00 244.9 F 12.8
16:00 257.5 F 19.6
18:00 268.8 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.5 F 29.3
2:00 271.4 F 28.7
3:00 272.5 F 29.4
4:00 272.0 F 28.8

Min 271.4 F 28.6
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 232.2 F 0.0
10:00 241.1 F 0.0
15:00 231.7 F 0.0
20:00 226.0 F 0.0
25:00 228.9 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

02/15/2023 12:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____