

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5294

Name: RITTER 1  
Sterilizer ID: V2178198  
01/29/2025 03:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.9 F 0.0  
2:00 158.5 F 3.0  
4:00 178.5 F 6.6  
6:00 202.9 F 10.6  
8:00 222.9 F 5.7  
10:00 223.7 F 8.4  
12:00 236.1 F 13.2  
14:00 236.3 F 9.8  
16:00 237.7 F 9.1  
18:00 247.6 F 14.2  
20:00 258.8 F 20.4  
22:00 269.1 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.6  
1:00 271.6 F 28.9  
2:00 272.5 F 29.3  
3:00 272.0 F 28.9  
4:00 271.3 F 28.5

Min 271.2 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.3 F 0.0  
5:00 212.7 F 0.0  
10:00 208.8 F 0.0  
15:00 228.8 F 0.0  
20:00 230.1 F 0.0  
25:00 226.5 F 0.0  
30:00 218.4 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 23:02  
STERILIZING: 4:00  
VENTING: 3:57  
DRYING: 30:00  
TOTAL CYCLE: 01:02:07  
CYCLE COMPLETE

01/29/2025 04:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_