

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1412

Name: RITTER 1  
Sterilizer ID: V2178198  
12/01/2021 08:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.8 F 0.0  
2:00 97.2 F 0.0  
4:00 129.1 F 0.0  
6:00 178.3 F 0.1  
8:00 196.8 F 0.2  
10:00 208.6 F 2.0  
12:00 219.3 F 5.0  
14:00 227.2 F 8.3  
16:00 234.8 F 11.9  
18:00 230.5 F 6.9  
20:00 241.0 F 11.8  
22:00 235.4 F 8.2  
24:00 242.9 F 11.7  
26:00 244.1 F 12.0  
28:00 255.8 F 18.0  
30:00 266.2 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 27.9  
1:00 272.2 F 28.9  
2:00 271.5 F 28.5  
3:00 271.2 F 28.3  
4:00 272.1 F 28.4  
  
Min 270.4 F 27.8  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.2 F 0.0  
5:00 235.9 F 0.0  
10:00 227.2 F 0.0  
15:00 226.2 F 0.0  
20:00 226.9 F 0.0  
25:00 226.9 F 0.0  
30:00 216.8 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 31:36  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:09:45  
CYCLE COMPLETE

12/01/2021 09:19 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02