
Midmark M11 - v1.0.5
Total Cycles: 2734

Name: RITTER2
Sterilizer ID: V2313228
08/04/2023 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.5 F 0.0
2:00 136.2 F 0.0
4:00 147.7 F 0.6
6:00 177.8 F 3.9
8:00 194.9 F 7.4
10:00 222.3 F 11.1
12:00 217.7 F 5.4
14:00 230.9 F 9.4
16:00 235.0 F 11.3
18:00 239.1 F 11.0
20:00 236.6 F 9.3
22:00 249.2 F 15.2
24:00 259.8 F 21.0
26:00 269.5 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.4 F 29.5
2:00 271.7 F 28.8
3:00 271.5 F 29.1
4:00 272.2 F 29.4

Min 271.1 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 188.4 F 0.0
10:00 189.7 F 0.0
15:00 187.9 F 0.0
20:00 185.4 F 0.0
25:00 183.5 F 0.0
30:00 182.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:06
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:05:31
CYCLE COMPLETE

08/04/2023 12:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____