
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001364
10:14 28/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:10
PRESSURIZING 4:11
STERILIZING 5:45
136.4 C 223kPa 5:45
136.4 C 224kPa 6:15
136.4 C 224kPa 6:45
136.3 C 224kPa 7:15
136.3 C 223kPa 7:45
136.3 C 222kPa 8:15
136.4 C 224kPa 8:45
136.2 C 221kPa 9:15
136.2 C 221kPa 9:15
VENTING 9:15
AIR DRYING 10:13
CYCLE COMPLETE 25:13

Digital Signature #
AFF197DF295E3B17
