
Midmark M11 - v1.0.5
Total Cycles: 2104

Name: RITTER2
Sterilizer ID: V2313228
02/28/2023 05:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.5 F 0.0
2:00 171.7 F 3.0
4:00 187.2 F 7.5
6:00 205.8 F 13.1
8:00 225.1 F 8.1
10:00 236.2 F 12.1
12:00 240.3 F 11.9
14:00 240.6 F 11.2
16:00 254.7 F 18.4
18:00 266.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.4 F 29.5
2:00 271.5 F 29.4
3:00 272.4 F 29.5
4:00 271.3 F 29.2

Min 271.3 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.7 F 0.0
5:00 235.5 F 0.0
10:00 224.6 F 0.0
15:00 231.4 F 0.0
20:00 225.7 F 0.0
25:00 221.1 F 0.0
30:00 208.8 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

02/28/2023 06:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____