
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000779
8:17 29/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:14
STERILIZING 3:18
136.0 C 228kPa 3:18
136.4 C 233kPa 3:48
136.2 C 232kPa 4:18
136.4 C 232kPa 4:48
136.3 C 232kPa 5:18
136.4 C 233kPa 5:48
136.3 C 233kPa 6:18
136.1 C 233kPa 6:48
136.1 C 233kPa 6:48
VENTING 6:48
AIR DRYING 7:39
CYCLE COMPLETE 7:41

Digital Signature #
27526038F467C1B9
