
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004298
10:59 06/09/2022
RUBBER AND PLASTICS
121 C for 15 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 4:00
STERILIZING 4:42
123.5 C 118kPa 4:42
123.4 C 118kPa 5:12
123.4 C 120kPa 5:42
123.2 C 118kPa 6:12
123.3 C 119kPa 6:42
123.0 C 117kPa 7:12
123.1 C 117kPa 7:42
123.3 C 116kPa 8:12
123.3 C 117kPa 8:42
123.4 C 118kPa 9:12
123.4 C 118kPa 9:42
123.4 C 119kPa 10:12
123.4 C 118kPa 10:42
123.4 C 119kPa 11:12
123.4 C 118kPa 11:42
123.4 C 118kPa 12:12
123.4 C 119kPa 12:42
123.4 C 118kPa 13:12
123.4 C 118kPa 13:42
123.3 C 119kPa 14:12
123.3 C 119kPa 14:42
123.3 C 118kPa 15:12
123.3 C 118kPa 15:42
123.3 C 118kPa 16:12
123.4 C 118kPa 16:42
123.4 C 118kPa 17:12

CYCLE ABORTED 17:21
STOP BUTTON PRESSED
NOT STERILE
END TIME 17:21

Digital Signature #
E8E9EF5995775C62
