
Midmark M11 - v1.0.5
Total Cycles: 2761

Name: RITTER2
Sterilizer ID: V2313228
08/14/2023 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.1 F 0.0
2:00 182.9 F 1.2
4:00 204.2 F 5.8
6:00 220.1 F 10.6
8:00 226.0 F 7.2
10:00 236.0 F 11.7
12:00 235.8 F 9.7
14:00 238.3 F 9.6
16:00 248.9 F 15.0
18:00 259.5 F 21.1
20:00 269.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 271.8 F 29.0
2:00 271.8 F 29.1
3:00 271.8 F 29.3
4:00 272.0 F 29.5

Min 271.2 F 28.8
Max 272.2 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 230.5 F 0.0
10:00 225.8 F 0.0
15:00 203.8 F 0.0
20:00 193.2 F 0.0
25:00 190.6 F 0.0
30:00 186.0 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

08/14/2023 03:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____