

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000144  
9:03 14/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:04  
PRESSURIZING 4:05  
STERILIZING 5:40  
136.0 C 222kPa 5:40  
136.4 C 223kPa 6:10  
136.4 C 226kPa 6:40  
136.4 C 224kPa 7:10  
136.4 C 225kPa 7:40  
136.1 C 224kPa 8:10  
136.4 C 224kPa 8:40  
136.4 C 224kPa 9:10  
136.4 C 224kPa 9:10  
VENTING 9:10  
AIR DRYING 10:11  
CYCLE COMPLETE 13:20  
-----

Digital Signature #  
24799045D75770F3  
-----