

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2198

Name: RITTER 1  
Sterilizer ID: V2178198  
08/04/2022 08:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.3 F 0.0  
2:00 162.8 F 0.9  
4:00 166.7 F 4.3  
6:00 188.3 F 8.4  
8:00 208.5 F 13.2  
10:00 223.9 F 6.5  
12:00 235.5 F 11.3  
14:00 232.9 F 7.6  
16:00 245.7 F 13.4  
18:00 242.7 F 11.7  
20:00 255.4 F 18.0  
22:00 266.3 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.3 F 28.7  
2:00 271.4 F 28.4  
3:00 272.1 F 28.6  
4:00 271.4 F 28.5

Min 271.1 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
5:00 234.0 F 0.0  
10:00 219.7 F 0.0  
15:00 226.6 F 0.0  
20:00 242.1 F 0.0  
25:00 233.0 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 23:37  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:39  
CYCLE COMPLETE

08/04/2022 09:30 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_