

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2963

Name: RITTER2  
Sterilizer ID: V2313228  
10/03/2023 10:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.9 F 0.0  
2:00 147.6 F 1.6  
4:00 161.5 F 5.5  
6:00 190.0 F 10.9  
8:00 215.5 F 5.3  
10:00 233.8 F 12.3  
12:00 238.0 F 10.6  
14:00 238.3 F 9.8  
16:00 253.8 F 17.5  
18:00 255.8 F 18.6  
20:00 269.7 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.6  
1:00 272.7 F 28.9  
2:00 273.0 F 29.3  
3:00 272.7 F 29.4  
4:00 271.9 F 28.7

Min 271.8 F 28.5  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 222.3 F 0.0  
5:00 213.3 F 0.0  
Dry Cycle Aborted

FILLING: 1:05  
HEATING: 20:42  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 5:60  
TOTAL CYCLE: 00:34:47

10/03/2023 11:02 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_