

-----  
Midmark M11 - v1.0.5  
Total Cycles: 7

Name: RITTER2  
Sterilizer ID: V2313228  
04/22/2021 06:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.1 F 0.0  
2:00 73.7 F 0.3  
4:00 77.2 F 1.3  
6:00 101.0 F 3.4  
8:00 128.7 F 6.2  
10:00 157.8 F 9.5  
12:00 181.7 F 13.0  
14:00 203.5 F 4.5  
16:00 215.1 F 8.2  
18:00 226.8 F 12.3  
20:00 230.4 F 9.0  
22:00 239.1 F 11.6  
24:00 243.8 F 13.2  
26:00 255.1 F 19.2  
28:00 265.5 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.5  
1:00 272.5 F 30.5  
2:00 272.2 F 29.7  
3:00 271.7 F 29.8  
4:00 272.6 F 30.4

Min 271.4 F 29.4  
Max 273.0 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.2 F 0.0  
5:00 213.1 F 0.0  
10:00 216.2 F 0.0  
15:00 206.9 F 0.0  
20:00 210.9 F 0.0  
25:00 205.6 F 0.0  
30:00 204.4 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 29:56  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:08:16  
CYCLE COMPLETE

04/22/2021 07:49 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_