
Midmark M11 - v1.0.0
Total Cycles: 4847

Name: RITTER 1
Sterilizer ID: V2178198
08/23/2024 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.7 F 0.0
2:00 173.1 F 0.8
4:00 177.7 F 2.8
6:00 197.7 F 7.2
8:00 216.4 F 11.8
10:00 222.6 F 6.3
12:00 234.3 F 11.4
14:00 233.7 F 8.1
16:00 243.5 F 11.6
18:00 245.0 F 12.6
20:00 257.2 F 19.0
22:00 268.1 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.4 F 28.7
2:00 271.4 F 28.0
3:00 272.5 F 28.8
4:00 271.6 F 28.2

Min 271.2 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 196.5 F 0.0
10:00 212.0 F 0.0
15:00 226.8 F 0.0
20:00 236.1 F 0.0
25:00 232.3 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 23:20
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:02:11
CYCLE COMPLETE

08/23/2024 10:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____