

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4943

Name: RITTER 1  
Sterilizer ID: V2178198  
09/27/2024 10:21 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.3 F 0.0  
2:00 73.1 F 0.0  
4:00 99.9 F 0.0  
6:00 131.4 F 0.0  
8:00 166.5 F 0.2  
10:00 191.0 F 0.4  
12:00 211.4 F 2.7  
14:00 218.3 F 4.2  
16:00 227.6 F 7.4  
18:00 237.1 F 11.2  
20:00 231.8 F 6.7  
22:00 245.3 F 13.0  
24:00 242.2 F 11.1  
26:00 238.6 F 9.5  
28:00 253.7 F 16.9  
30:00 266.1 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.9  
1:00 272.6 F 28.4  
2:00 272.2 F 28.4  
3:00 272.3 F 28.1

Min 271.6 F 27.7  
Max 272.9 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 163.3 F 0.0  
5:00 182.0 F 0.0  
10:00 172.0 F 0.0  
15:00 169.9 F 0.0  
20:00 169.8 F 0.0  
25:00 170.4 F 0.0  
30:00 176.8 F 0.0  
Drying Complete

FILLING: 0:43  
HEATING: 31:34  
STERILIZING: 3:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:08:21  
CYCLE COMPLETE

09/27/2024 11:29 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_