
Midmark M11 - v1.0.0
Total Cycles: 4068

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2023 05:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	165.5 F	0.0
2:00	160.9 F	2.9
4:00	172.5 F	6.2
6:00	190.5 F	10.0
8:00	217.5 F	7.6
10:00	223.8 F	7.6
12:00	236.7 F	12.0
14:00	232.6 F	8.4
16:00	244.3 F	12.6
18:00	243.2 F	12.0
20:00	254.7 F	18.0
22:00	265.0 F	24.2
24:00	271.2 F	28.5

STERILIZING

mm:ss	Degrees	PSI
0:00	271.2 F	28.5
1:00	272.4 F	29.3
2:00	272.2 F	28.9
3:00	271.2 F	28.4
4:00	271.6 F	28.9

Min 271.0 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	206.5 F	0.0
	5:00	229.0 F	0.0
	10:00	212.1 F	0.0
	15:00	216.2 F	0.0
	20:00	213.6 F	0.0
	25:00	209.7 F	0.0
Drying Complete	30:00	208.4 F	0.0

FILLING: 1:25
HEATING: 24:00
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:09
CYCLE COMPLETE

11/29/2023 06:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____