
Midmark M11 - v1.0.5
Total Cycles: 4349

Name: RITTER2
Sterilizer ID: V2313228
02/11/2025 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.3 F 0.0
2:00 69.4 F 0.0
4:00 71.8 F 0.4
6:00 89.3 F 2.3
8:00 124.5 F 4.9
10:00 153.0 F 7.5
12:00 170.7 F 10.6
14:00 204.2 F 10.6
16:00 207.3 F 4.4
18:00 216.7 F 7.4
20:00 226.6 F 10.7
22:00 231.5 F 8.7
24:00 238.4 F 10.7
26:00 236.3 F 8.8
28:00 246.5 F 13.7
30:00 256.6 F 19.2
32:00 266.0 F 25.2
34:00 271.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.7
1:00 272.0 F 30.0
2:00 271.9 F 30.1
3:00 272.0 F 30.1
4:00 271.9 F 30.1

Min 271.6 F 29.6
Max 272.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 204.8 F 0.0
10:00 224.6 F 0.0
15:00 221.0 F 0.0
20:00 182.2 F 0.0
25:00 178.6 F 0.0
30:00 177.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 34:32
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:13:15
CYCLE COMPLETE

02/11/2025 12:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____