
Midmark M11 - v1.0.5
Total Cycles: 3876

Name: RITTER2
Sterilizer ID: V2313228
08/20/2024 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.8 F 0.0
2:00 100.7 F 0.0
4:00 120.0 F 0.2
6:00 155.3 F 2.5
8:00 180.4 F 6.3
10:00 203.1 F 10.8
12:00 213.8 F 5.0
14:00 223.5 F 8.4
16:00 235.0 F 13.0
18:00 235.8 F 9.9
20:00 236.8 F 9.1
22:00 247.0 F 14.2
24:00 258.3 F 20.4
26:00 269.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.6
2:00 271.5 F 28.9
3:00 272.7 F 29.8
4:00 272.1 F 29.2

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 206.0 F 0.0
10:00 214.3 F 0.0
15:00 214.9 F 0.0
20:00 213.3 F 0.0
25:00 211.2 F 0.0
30:00 207.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 27:09
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:05:30
CYCLE COMPLETE

08/20/2024 12:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____