
Midmark M11 - v1.0.5
Total Cycles: 2372

Name: RITTER2
Sterilizer ID: V2313228
05/16/2023 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.2 F 0.0
2:00 154.8 F 0.2
4:00 167.8 F 1.1
6:00 190.2 F 5.8
8:00 211.3 F 10.6
10:00 219.2 F 5.6
12:00 233.5 F 11.3
14:00 233.4 F 8.8
16:00 238.8 F 10.2
18:00 247.2 F 14.4
20:00 259.6 F 21.3
22:00 270.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.5 F 29.6
2:00 271.8 F 29.5
3:00 272.3 F 29.5
4:00 271.7 F 29.4

Min 271.5 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 201.6 F 0.0
10:00 217.7 F 0.0
15:00 224.2 F 0.0
20:00 222.9 F 0.0
25:00 224.4 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:46
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:58
CYCLE COMPLETE

05/16/2023 09:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____