

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000129  
11:54 08/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:28  
PRESSURIZING 4:28  
STERILIZING 6:02  
136.5 C 225kPa 6:02  
136.4 C 225kPa 6:32  
136.3 C 224kPa 7:02  
136.4 C 223kPa 7:32  
136.4 C 225kPa 8:02  
136.3 C 224kPa 8:32  
136.4 C 225kPa 9:02  
136.4 C 225kPa 9:32  
136.4 C 225kPa 9:32  
VENTING 9:32  
AIR DRYING 10:37  
CYCLE COMPLETE 14:53  
-----

Digital Signature #  
136B154ACDEBAA34  
-----