

-----  
Midmark M11 - v1.0.0  
Total Cycles: 741

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2021 10:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.6 F 0.0  
2:00 145.4 F 1.0  
4:00 195.3 F 3.8  
6:00 189.7 F 8.4  
8:00 221.2 F 9.3  
10:00 222.7 F 8.6  
12:00 234.8 F 10.6  
14:00 239.3 F 11.6  
16:00 239.2 F 9.8  
18:00 252.7 F 16.5  
20:00 264.4 F 23.4  
22:00 271.6 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.4 F 29.8  
2:00 271.9 F 29.1  
3:00 271.7 F 29.2  
4:00 272.2 F 29.8

Min 271.0 F 28.8  
Max 273.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.2 F 0.0  
5:00 233.6 F 0.0  
10:00 229.9 F 0.0  
15:00 228.0 F 0.0  
20:00 229.3 F 0.0  
25:00 230.2 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 22:10  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 01:00:05  
CYCLE COMPLETE

03/24/2021 11:56 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01