
Midmark M11 - v1.0.0
Total Cycles: 664

Name: MIDMARK
Sterilizer ID: V2178198
02/11/2021 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

C232: FILL MODE
WATER LOW
CHAMBER TEMP 71F
CHAMBER PRESS 0.0PSI

** ITEMS NOT STERILE **

02/11/2021 09:18 AM