
Midmark M11 - v1.0.0
Total Cycles: 3949

Name: RITTER 1
Sterilizer ID: V2178198
10/27/2023 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.2 F 0.0
2:00 124.6 F 1.1
4:00 133.1 F 3.4
6:00 158.4 F 6.7
8:00 182.8 F 10.9
10:00 208.3 F 3.2
12:00 217.4 F 7.3
14:00 229.2 F 11.6
16:00 230.9 F 7.6
18:00 243.1 F 13.0
20:00 241.0 F 11.0
22:00 253.2 F 17.1
24:00 264.1 F 23.5
26:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 28.9
2:00 271.6 F 28.6
3:00 272.3 F 28.9
4:00 271.6 F 28.6

Min 271.4 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 192.5 F 0.0
10:00 206.2 F 0.0
15:00 208.0 F 0.0
20:00 207.9 F 0.0
25:00 210.6 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 26:04
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

10/27/2023 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____