
Midmark M11 - v1.0.5
Total Cycles: 33

Name: RITTER2
Sterilizer ID: V2313228
05/19/2021 02:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.0 F 0.0
2:00 162.1 F 3.2
4:00 184.0 F 7.1
6:00 203.9 F 11.9
8:00 221.2 F 6.5
10:00 234.0 F 12.1
12:00 236.4 F 9.9
14:00 237.1 F 9.1
16:00 250.5 F 15.9
18:00 262.6 F 22.9
20:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.2
2:00 271.3 F 28.9
3:00 272.5 F 29.3
4:00 271.3 F 28.6

Min 271.1 F 28.6
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 230.5 F 0.0
10:00 227.5 F 0.0
15:00 223.1 F 0.0
20:00 210.9 F 0.0
25:00 214.1 F 0.0
30:00 206.6 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:10
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

05/19/2021 03:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____