
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007522
9:35 10/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:51
STERILIZING 4:03
136.4 C 222kPa 4:03
136.4 C 225kPa 4:33
136.4 C 222kPa 5:03
136.4 C 224kPa 5:33
136.3 C 223kPa 6:03
136.3 C 224kPa 6:33
136.4 C 222kPa 7:03
136.3 C 222kPa 7:33
136.3 C 222kPa 7:33
VENTING 7:33
AIR DRYING 8:25
CYCLE COMPLETE 38:25

Digital Signature #
EAAF0D566B904667
