
Midmark M11 - v1.0.5
Total Cycles: 3458

Name: RITTER2
Sterilizer ID: V2313228
03/15/2024 02:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.6 F 0.0
2:00 118.6 F 0.3
4:00 141.9 F 2.4
6:00 166.4 F 6.0
8:00 191.5 F 10.8
10:00 212.4 F 4.1
12:00 221.7 F 9.6
14:00 231.0 F 9.1
16:00 240.4 F 12.0
18:00 239.1 F 10.6
20:00 252.0 F 17.2
22:00 263.8 F 24.2
24:00 271.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.8 F 30.3
2:00 272.0 F 29.6
3:00 271.1 F 29.3
4:00 272.1 F 30.3

Min 271.1 F 29.2
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 221.4 F 0.0
10:00 190.9 F 0.0
15:00 184.6 F 0.0
20:00 182.3 F 0.0
25:00 180.6 F 0.0
30:00 179.0 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 24:05
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:20
CYCLE COMPLETE

03/15/2024 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____