

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2448

Name: RITTER2  
Sterilizer ID: V2313228  
06/01/2023 07:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.4 F 0.0  
2:00 74.7 F 0.0  
4:00 79.0 F 0.0  
6:00 103.1 F 0.0  
8:00 150.7 F 1.6  
10:00 166.6 F 5.0  
12:00 206.9 F 8.8  
14:00 215.5 F 12.8  
16:00 219.6 F 6.4  
18:00 230.3 F 10.6  
20:00 232.0 F 8.1  
22:00 241.2 F 12.4  
24:00 239.5 F 10.6  
26:00 251.8 F 16.7  
28:00 262.3 F 22.9  
30:00 272.3 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.5 F 29.7  
2:00 271.5 F 28.9  
3:00 271.8 F 29.7  
4:00 272.2 F 29.4

Min 271.0 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.2 F 0.0  
5:00 218.7 F 0.0  
10:00 226.0 F 0.0  
15:00 228.4 F 0.0  
20:00 226.5 F 0.0  
25:00 221.1 F 0.0  
30:00 215.1 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 30:30  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:25  
CYCLE COMPLETE

06/01/2023 08:21 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_